

(Alco - Sensor
FST

701

~~Weekly~~ ^{Monthly} Infrared Instrument
Calibration Verification Test Record

Department: Cazenovia Police Dept Instrument Serial #: 038780

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
10/2/19	19180/792	33.97			1983	.102			W Stevens	
10/31/19	19180/670	33.96			1983	.105			W Stevens	
11/30/19	19180/137	33.96			1983	.099			W Stevens	

* - Please enter a √ to denote simulator solution change taking place

Alco-Sensor
FST

701

Monthly
Weekly Infrared Instrument
Calibration Verification Test Record

Department: Cazenovia Police Dept.

Instrument Serial #: 038780

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
10/5/18	18170/177	34.00			M010621	.101			W. Stevens	
11/1/18	18230/990	34.2			M010621	.102			W. Stevens	
12/1/18	18230/1022	34.2			M010621	.102			M. Schwarz	✓
12/29/18	18310/681	34.1			M010621	.103			W. Stevens	
1/31/19	18310/137	34.1			M010621	.108			W. Stevens	✓
2/28/19	18360/255	34.00			M010621	.108			W. Stevens	
4/1/19	18360/302	34.2			M010621	.105			W. Stevens	
4/30/19	19030/1350	34.2			M010621	.097			W. Stevens	
6/1/19	19030/1352	34.00			M010621	.100			W. Stevens	
6/29/19	19090/546	34.1			M010621	.101			W. Stevens	
8/8/19	19090/549	34.1			M010621	.099			W. Stevens	
8/30/19	19130/745	34.1			M010621	.104			W. Stevens	

* - Please enter a ✓ to denote simulator solution change taking place

Alco - Sensor
FST

701

Monthly Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia Police Dept Instrument Serial #: 038780

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
9/30/17	17150/1171	34.00			M 010621	.102			W. Stevens	
10/31/17	17220/909	34.00			M010621	.102			W. Stevens	
12/1/17	17220/882	34.00			M010621	.101			W. Stevens	
12/30/17	17310/638	34.00			M010621	.103			W. Stevens	
1/31/18	17310/202	34.00			M010621	.106			W. Stevens	
3/1/18	17390/269	34.00			M010621	.106			W. Stevens	
4/7/18	17390/277	34.00			M010621	.104			W. Stevens	
5/1/18	18040/247	34.00			M010621	.096			W. Stevens	
5/31/18	18040/19	34.00			M010621	.100			W. Stevens	
6/30/18	18100/448	34.00			M010621	.100			W. Stevens	
8/2/18	18100/459	34.00			M010621	.099			W. Stevens	
8/31/18	18170/245	34.00 34.2			M010621	.099			W. Stevens	

PBT in vehicle and vehicle was on repair

* - Please enter a √ to denote simulator solution change taking place