Preliminary Breath Tester
Monthly Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139406

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Officer Performing Verification	Sim. Sol. Change *
11/14/19	19180/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
11/21/19	19180/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
11/27/19	19250/881	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	solution Change
12/5/19	19250/881	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
12/12/19	19250/881	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

Preliminary Breath Tester
Monthly Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Officer Performing Verification	Sim. Sol. Change *
8/29/19	19130/793	34.01	34.01	34.00	MP4614	.10	.10	.10	W Stevens	Bottle Change
9/5/19	19130/793	34.00	34.00	34.01	MP4614	.10	.10	.10	W Stevens	
9/12/19	19130/793	34.01	34.01	34.00	MP4614	.10	.10	.10	W Stevens	
9/18/19	19130/793	34.00	34.00	34.00	MP4614	.10	.10	.10	W Stevens	
9/25/19	out for calibration				MP4614	–	–	-	W. Stevens	WS ~~solution change~~
9/26/19	19180/670	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	Solution change
10/3/19	19180/670	34.00	34.00	34.00	MP4614	.10	.10	.10	W Stevens	
10/10/19	19180/670	34.00	34.00	33.99	MP4614	.10	.10	.10	W. Stevens	
10/17/19	19180/670	34.00	34.00	33.99	MP4614	.09	.10	.10	W. Stevens	
10/24/19	19180/670	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
10/31/19	19180/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	Bottle Change
11/7/19	19180/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

Preliminary Breath Tester
Monthly Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Officer Performing Verification	Sim. Sol. Change *
6/6/19	19090/546	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
6/14/19	19090/546	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
6/20/19	19090/546	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
6/27/19	19090/549	34.01	34.00	34.00	MP4614	.10	.10	.10	W Stevens	Bottle Change
7/4/19	19090/549	34.00	34.01	34.00	MP4614	.10	.10	.10	W Stevens	
7/11/19	19090/549	34.00	34.00	34.01	MP4614	.10	.10	.10	W. Stevens	
7/18/19	19090/549	34.00	34.00	34.00	MP4614	.10	.10	.10	W Stevens	
7/25/19	19090/549	34.01	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
8/1/19	19130/745	34.00	34.01	34.00	MP4614	.10	.10	.10	W. Stevens	solution Change
8/8/19	19130/745	34.00	34.01	34.01	MP4614	.10	.10	.10	W. Stevens	
8/16/19	19130/745	34.01	34.01	34.00	MP4614	.10	.10	.10	W. Stevens	
8/22/19	19130/745	34.01	34.00	34.01	MP4614	.09	.10	.10	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
3/14/19	18360/302	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
3/20/19	18360/302	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
3/28/19	19030/1350	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	solution change
4/4/19	19030/1350	34.00	34.00	34.00	MP4619	.10	.10	.10	W. Stevens	
4/11/19	19030/1350	34.00	34.00	34.00	MP4619	.10	.10	.10	W. Stevens	
4/18/19	19030/1350	34.00	34.00	34.00	MP4619	.10	.10	.10	W. Stevens	
4/25/19	19030/1352	34.00	34.00	34.00	MP4619	.10	.10	.10	W. Stevens	Bottle Change
5/2/19	19030/1352	34.00	34.00	34.00	MP4619	.10	.10	.10	W. Stevens	
5/9/19	19030/1352	34.00	34.00	34.00	MP4619	.10	.10	.10	W. Stevens	
5/16/19	19030/1352	34.00	34.00	34.00	MP4619	.10	.10	.10	Schwock	
05/23/19	19030/1352	34.00	34.00	34.00	MP4619	.10	.10	.10	Schwock	
6/1/19	19690/546	34.00	34.00	34.00	MP619	.10	.10	.10	W. Stevens	solution change

Did not print 1st Time

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
12/20/18	18310/681	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
12/27/18	18310/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	Bottle Change
1/3/19	18310/137	34.00	34.00	34.00	MP4614	.10	.10	.10	~~W. St~~ as W Stevens	
1/10/19	18310/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W Stevens	
1/17/19	18310/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
1/24/19	18310/137	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
1/31/19	18360/255	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	solution Change
2/7/19	18360/255	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
2/14/19	18360/255	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
2/21/19	18360/255	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
2/28/19	18360/302	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	Bottle Change
3/7/19	18360/302	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
9/28/18	18230/990	34.00	34.00	34.00	~~1176~~ WS MP4614	.10	.10	.10	weaver/MCSO	SOL. change
10/4/18	18230/990	34.00	34.00	34.00	~~1176~~ WS MP4614	.10	.10	.10	W. Stevens	
10/11/18	18230/990	34.00	34.00	34.00	~~1176~~ WS MP4614	.10	.10	.10	W. Stevens	
10/18/18	18230/990	34.00	34.00	34.00	~~1176~~ WS MP4614	.10	.10	.10	W. Stevens	
10/25/18	18230/990	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
11/1/18	18230/1022	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	Bottle Change
11/8/18	18230/1022	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
11/15/18	18230/1022	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
11/21/18	18230/1022	34.00	34.00	34.00	MP4614	.10	.10	.10	W. Stevens	
12/1/18	18310/681	34.00	34.00	34.00	MP4614	.10	.10	.10	M. Schwab	Sol Change
12/7/18	18310/681	34.00	34.00	34.00	MP4614	.10	.10	.10	M. Schwab	
12/13/18	18310/681	34.00	34.00	34.00	MP4614	.10	.10	.10	W Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

(

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
7/5/18	18100/459	34.01	34.01	34.00	1176	.09	.09	.09	W. Stevens	
7/12/18	18100/459	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
7/19/18	18100/459	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
7/27/18	18100/459	34.00	34.01	34.01	1176	.09	.09	.09	W. Stevens	
8/2/18	18170/245	34.01	34.00	34.00	1176	.09	.09	.09	W. Stevens	solution Change
8/9/18	18170/245	34.01	34.01	33.99	1176	.09	.09	.09	W. Stevens	
8/15/18	18170/245	34.00	34.00	33.99	1176	.09	.09	.09	W. Stevens	
8/23/18	18170/245	34.00	34.01	34.00	1176	.09	.09	.09	W. Stevens	
8/30/18	18170/177	34.01	34.02	34.01	1176	.09	.09	.09	W. Stevens	Bottle Change
9/6/18	18170/177	34.00	34.00	33.99	1176	.09	.09	.09	W. Stevens	
9/13/18	18170/177	33.99	34.01	33.99	1176	.09	.09	.09	W. Stevens	
9/20/18	18170/177	34.00	34.01	33.99	1176	.09	.09	.09	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

(

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
4/5/18	18040/247	33.98	33.99	34.00	1176	.09	.09	.09	W. Stevens	
4/12/18	18040/247	33.99	34.00	34.00	1176	.09	.09	.09	W. Stevens	
4/19/18	18040/247	33.98	34.00	33.99	1176	.09	.09	.09	W. Stevens	
4/26/18	18040/19	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	Bottle change
5/3/18	18040/19	34.01	34.01	34.00	1176	.09	.09	.09	W. Stevens	
5/10/18	18040/19	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
5/25/18	18040/19	34.00	34.00	34.00	1176	.09	.09	.09	M White	
5/31/18	18100/448	34.00	34.00	34.00	1176	.09	.09	.10	W. Stevens	Solution Change
6/7/18	18100/448	33.98	34.00	34.00	1176	.09	.09	.09	W. Stevens	
6/14/18	18100/448	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
6/21/18	18100/448	34.00	34.00	33.99	1176	.09	.09	.09	W. Stevens	
6/28/18	18100/459	34.00	34.00	34.01	1176	.09	.09	.09	W. Stevens	Bottle Change

*** - Please enter a √ to denote simulator solution change taking place**

(

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
1/11/18	17310 / 202	34.00	33.99	33.98	1176	.09	.09	.09	W. Stevens	
1/18/18	17310 / 202	34.00	34.00	33.99	1176	.09	.09	.09	W. Stevens	
1/25/18	17310 / 202	33.96	33.98	33.99	1176	.09	.09	.09	W. Stevens	
1/31/18	17390/269	33.99	33.98	33.98	1176	.09	.09	.09	W. Stevens	Solution Change
2/8/18	17390/269	33.98	34.00	34.00	1176	.09	.09	.09	W. Stevens	
2/15/18	17390/269	34.00	33.99	34.00	1176	.09	.09	.09	W. Stevens	
2/22/18	17390/269	34.01	34.00	34.00	1176	.09	.09	.09	W. Stevens	
3/1/18	17390/277	34.00	34.00	33.99	1176	.09	.09	.09	W. Stevens	Bottle Change
3/10/18	17390 / 277	33.98	33.98	33.99	1176	.09	.09	.09	W. Stevens	
3/15/18	17390/277	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
3/22/18	17390/277	33.99	34.00	34.00	1176	.09	.09	.09	W. Stevens	
3/29/18	18040/274	33.99	33.99	33.99	1176	.09	.09	.09	W. Stevens	solution Change

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
10/19/17	17220/909	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
10/27/17	17220/882	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	Bottle Change
11/2/17	17220/882	34.00	34.00	33.99	1176	.09	.09	.09	W. Stevens	
11/9/17	17220/882	33.99	33.99	34.00	1176	.09	.09	.09	W. Stevens	
11/16/17	17220/882	33.98	34.00	33.99	1176	.09	.09	.09	W. Stevens	
11/22/17	17220/882	33.98	33.99	33.99	1176	.09	.09	.09	W. Stevens	
12/1/17	17310/638	33.98	33.98	33.97	1176	.09	.09	.09	W. Stevens	Solution Change
12/7/17	17310/638	33.98	33.99	33.98	1176	.09	.09	.09	W. Stevens	
12/14/17	17310/638	33.99	33.99	33.99	1176	.09	.09	.09	W. Stevens	
12/21/17	17310/638	34.00	33.98	33.98	1176	.09	.09	.09	W. Stevens	
12/28/17	17310/202	33.99	33.99	33.99	1176	.09	.09	.09	W. Stevens	Bottle Change
1/4/18	17310/202	33.99	34.01	34.00	1176	.09	.09	.09	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

(

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
7/27/17	17150/1385	34.01	34.00	34.00	1176	.10	.10	.10	W. Stevens	Solution change
8/3/17	17150/1385	34.01	34.00	34.01	1176	.10	.10	.10	W. Stevens	
8/10/17	17150/1385	34.01	34.02	33.99	1176	.10	.10	.10	W. Stevens	
8/17/17	17150/1385	34.01	34.00	34.00	1176	.10	.10	.10	W. Stevens	
8/24/17	17150/1385	34.00	33.99	34.01	1176	.10	.10	.10	W. Stevens	
8/31/17	17150/1171	33.99	34.01	34.00	1176	.10	.10	.10	W. Stevens	Bottle change
calibration 9/5/17 → 9/7/17	17150/1171	33.97	34.00	34.00	1176	.09	.09	.09	W. Stevens	
9/14/17	17150/1171	34.00	34.01	34.01	1176	.09	.09	.09	W. Stevens	
9/21/17	17150/1171	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
9/29/17	17220/909	34.01	34.01	33.99	1176	.09	.09	.09	W. Stevens	Solution change
10/5/17	17220/909	34.00	34.02	34.01	1176	.09	.09	.09	W. Stevens	
10/12/17	17220/909	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
5/5/17	17040/142	33.98	33.99	33.99	1176	.09	.09	.10	W. Stevens	
5/12/17	17040/142	34.01	34.00	33.98	1176	.09	.09	.10	W. Stevens	
5/18/17	17040	34.01	34.01	34.00	1176	.09	.10	.10	M. White	
5/26/17	17040	33.98	34.00	34.01	1176	.09	.09	.09	M. White	
6/1/17	~~17040~~ 17120/615	34.00	34.01	33.98	1176	.10	.10	.10	W. Stevens	Bottle Solution Change ✓
6/8/17	17120/615	33.99	34.00	33.99	1176	.10	.10	.10	W. Stevens	
6/14/17	17120/615	34.00	34.00	34.00	1176	.10	.10	.10	W. Stevens	
6/22/17	17120/615	34.00	34.00	34.00	1176	.10	.10	.10	W. Stevens	
6/30/17	17120/668	34.00	33.99	34.00	1176	.10	.10	.10	W. Stevens	Bottle Change
7/6/17	17120/668	34.00	34.00	34.00	1176	.10	.10	.10	W. Stevens	
7/13/17	17120/668	34.01	34.00	34.01	1176	.10	.10	.10	W. Stevens	
7/20/17	17120/668	34.00	34.00	34.00	1176	.10	.10	.10	W Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Cazenovia P.D. **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
1/26/17	16380/772	33.98	34.00	33.99	1176	.09	.10	.10	W. Stevens	Solution Change ✓
2/3/17	16380/772	34.00	33.99	33.99	1176	.09	.10	.10	W. Stevens	
2/9/17	16380/772	34.00	34.00	33.99	1176	.09	.09	.10	W. Stevens	
2/16/17	16380/804	34.00	34.00	33.99	1176	.09	.09	.10	W. Stevens	Bottle Change
2/23/17	16380/804	34.00	33.99	33.99	1176	.09	.10	.10	W. Stevens	
3/3/17	16380/804	33.98	33.99	34.00	1176	.09	.09	.09	W. Stevens	
3/23/17	16380/804	33.99	34.01	34.00	1176	.09	.09	.09	W. Stevens	
3/30/17	17040/89	33.98	33.99	33.99	1176	.09	.10	.10	W. Stevens	Solution Change ✓
4/6/17	17040/89	33.98	34.01	34.00	1176	.09	.10	.10	W. Stevens	
4/13/17	17040/89	33.99	33.99	33.99	1176	.09	.10	.10	W. Stevens	
4/21/17	17040/89	33.99	33.98	33.99	1176	.09	.10	.10	W. Stevens	
4/28/17	17040/142	33.98	33.99	33.98	1176	.10	.10	.10	W. Stevens	Bottle Change

INC

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: CAZENOVIA PD **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
10/07 2016	16170/243	33.99	34.00	34.00	1176	.09	.09	.09	JP WATKINS	
10/16 2016	16170/243	34.00	34.00	33.99	1176	.09	.09	.09	JP WATKINS	
10/22 2016	16170/243	33.99	33.99	33.98	1176	.09	.09	.09	JP WATKINS	
11/01 2016	16170/307	33.98	34.00	33.99	1176	.09	.09	.10	JP WATKINS	BOTTLE CHANGE
11/07 2016	16170/307	33.98	34.00	33.98	1176	.09	.09	.09	JP WATKINS	
11/14 2016	16170/307	34.00	33.99	33.99	1176	.09	.09	.09	JP WATKINS	
12/05 2016	16240/956	34.00	33.98	33.98	1176	.09	.09	.09	JP WATKINS	SOLUTION CHANGE
12/15 2016	16240/956	34.00	34.00	34.00	1176	.09	.09	.09	W. Stevens	
12/29 2016	16240/956	34.00	33.98	34.00	1176	.09	.09	.09	W. Stevens	
1/5/17	16240/953	34.00	~~33.98~~ 33.98	33.99	1176	.09	.09	.09	W. Stevens	Bottle Change
1/12/17	16240/953	33.98	34.00	33.98	1176	.09	.09	.09	W. Stevens	
1/19/17	16240/953	34.00	34.00	33.99	1176	.09	.09	.09	W. Stevens	

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: CAZENOVIA PD **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
7/09 2016	16090/671	33.99	34.00	33.99	1176	.09	.10	.10	JPWATKINS	
7/15 2016	16090/671	34.01	34.00	33.99	1176	.09	.09	.10	JPWATKINS	
7/22 2016	16090/671	34.01	34.00	34.00	1176	.09	.09	.09	JPWATKINS	
7/29 2016	16120/345	34.00	33.99	34.00	1176	.09	.09	.09	JPWATKINS	✓ SOLUTION CHANGE
8/06 2016	16120/345	34.01	34.00	34.01	1176	.09	.09	.09	JPWATKINS	
8/12 2016	16120/345	34.00	33.99	34.02	1176	.09	.09	.09	JPWATKINS	
8/20 2016	16120/345	34.00	33.99	34.01	1176	.09	.09	.09	JPWATKINS	
8/28 2016	16120/298	34.00	34.00	34.01	1176	.09	.09	.09	JPWATKINS	BOTTLE CHANGE
9/02 2016	16120/298	34.00	34.00	34.00	1176	.09	.09	.09	JPWATKINS	
9/22 2016	16170/243	34.00	34.01	34.00	1176	.09	.09	INTERFERENCE	JPWATKINS	✓ SOLUTION CHANGE
9/22 2016	16170/243	33.99	34.00	33.99	1176	.09	.09	.10	JPWATKINS	
9/30 2016	16170/243	33.99	33.99	34.00	1176	.09	.09	.09	JPWATKINS	

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: CAZENOVIA PD **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
4/23 2016	16050/616	33.99	34.00	34.01	1176	.09	.09	.09	JP WATKINS	
4/30 2016	16050/613	34.00	33.96	33.98	1176	.09	.10	.10	JP WATKINS	BOTTLE CHANGE
5/5 2016	16050/613	34.00	34.00	33.99	1176	.09	.09	.10	JP WATKINS	
5/14 2016	16050/613	34.01	33.99	33.98	1176	.09	.09	.09	JP WATKINS	
5/21 2016	16050/613	33.99	33.99	34.00	1176	.09	.09	.09	JP WATKINS	
5/28 2016	16050/613	34.01	33.99	34.01	1176	.09	.09	.09	JP WATKINS	
5/28 2016	16090/680	34.00	34.00	34.01	1176	.10	.10	.10	JP WATKINS	✓ SOLUTION CHANGE
6/4 2016	16090/680	34.00	34.00	33.99	1176	.09	.09	.10	JP WATKINS	
6/10 2016	16090/680	34.00	34.01	34.00	1176	.09	.09	.09	JP WATKINS	
6/17 2016	16090/680	34.00	34.00	34.00	1176	.09	.09	.09	JP WATKINS	
6/24 2016	16090/680	34.01	33.99	34.02	1176	.09	.09	.09	JP WATKINS	
7/03 2016	16090/671	34.01	34.00	34.00	1176	.09	.10	.10	JP WATKINS	BOTTLE CHANGE

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: CAZENOVIA PD **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
2/06 2016	15250/855	33.99	34.01	33.99	1176	.09	.09	.10	JPWATKINS	
2/13 2016	15250/855	33.99	33.99	33.98	1176	.09	.09	.10	JPWATKINS	
2/19 2016	15250/855	33.98	34.00	34.00	1176	.09	.09	.09	JPWATKINS	
2/26 2016	15250/855	34.00	34.00	33.99	1176	.09	.10	.09	JPWATKINS	
2/27 2016	15250/613	33.99	33.99	33.99	1176	.09	.09	.10	JPWATKINS	BOTTLE CHANGE
3/05 2016	15250/613	34.00	33.99	33.98	1176	.09	.09	.09	JPWATKINS	
3/12 2016	15250/613	34.01	34.00	33.99	1176	.09	.09	.09	JPWATKINS	
3/19 2016	15250/613	34.01	33.99	33.99	1176	.09	.09	.09	JPWATKINS	
3/26 2016	16050/616	34.00	33.99	33.99	1176	.09	.09	.09	JPWATKINS	✓ SOL CH.
4/03 2016	16050/616	33.99	33.99	34.00	1176	.09	.09	.09	JPWATKINS	
4/09 2016	16050/616	34.00	34.01	33.99	1176	.09	.09	.09	JPWATKINS	
4/17 2016	16050/616	34.01	34.00	33.98	1176	.09	.09	.09	JPWATKINS	

*** - Please enter a √ to denote simulator solution change taking place**

Infrared Instrument
Weekly Calibration Verification Test Record

Department: CAZENOVIA PD **Instrument Serial #:** 139906

Test Date	Simulator Solution Lot #	Simulator Temperature 1	2	3	Sim. Serial#	Test Results 1	2	3	Breath Analysis Supervisor	Sim. Sol. Change *
11/13 2015	15180/33	34.00	34.00	34.00	1176	.09	.09	.09	JP WATKINS	
11/19 2015	15180/33	34.01	34.00	33.99	1176	.09	.09	.09	JP WATKINS	
11/25 2015	15180/33	33.98	34.01	33.99	1176	.09	.09	.09	JP WATKINS	
12/03 2015	15210/79	33.98	34.00	33.99	1176	.09	.09	.10	JP WATKINS	✓ SOL CH.
12/10 2015	15210/79	34.01	34.00	33.99	1176	.09	.09	.10	JP WATKINS	
12/17 2015	15210/79	34.00	34.00	33.99	1176	.09	.09	.09	JP WATKINS	
12/23 2015	15210/79	34.01	33.98	34.00	1176	.09	.09	.09	JP WATKINS	
1/1 2016	15210/247	34.00	33.99	34.00	1176	.09	.09	.10	JP WATKINS	BOTTLE CHANGE
1/8 2016	15210/247	34.00	34.00	33.99	1176	.09	.09	.10	JP WATKINS	
1/16 2016	15210/247	34.00	33.98	33.99	1176	.09	.09	.09	JP WATKINS	
1/22 2016	15210/247	33.99	33.99	33.99	1176	.09	.09	.09	JP WATKINS	
1/30 2016	15250/855	34.00	33.99	33.98	1176	.09	.10	.10	JP WATKINS	✓ SOL CH

*** - Please enter a √ to denote simulator solution change taking place**