

Infrared Instrument
Weekly Calibration Verification Test Record

Department: MC50

Instrument Serial #: 136506

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
2/16/19	18360	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
2/23/19	18360	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
3/1/19	18360	34.00	34.00	34.00	53030	.09	.10	.10	Weaver	✓
3/8/19	18360	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
3/15/19	18360	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
3/22/19	18360	34.00	34.00	34.00	53030	.09	.09	.10	Weaver	
4/1/19	19030	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	✓
4/8/19	19030	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
4/14/19	19030	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
4/21/19	19030	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
5/2/19	19030	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	✓
5/9/19	19030	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	

* - Please enter a ✓ to denote simulator solution change taking place