

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: MCSO

Instrument Serial #: 136506

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
5/9/19	19030	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
6/30/19	19090	33.99	33.99	33.99	53030	.10	.10	.10	Weaver	✓
7/14/19	19090	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
7/21/19	19090	34.00	34.00	34.00	53030	.09	.10	.10	Weaver	
7/28/19	19090	Blank			Test Error				Weaver	
7/28/19	19090	34.00	33.99	33.99	35030	.09	.09	.09	Weaver	
8/2/19	19130	33.99	33.99	33.99	35030	.09	.10	.10	Weaver	✓
8/12/19	19130	33.99	34.00	34.00	35030	.09	.10	.10	Weaver	
8/18/19	19130	34.00	34.00	34.00	35030	.09	.09	.10	Weaver	
8/25/19	19130	33.99	34.00	34.00	35030	.09	.09	.09	Weaver	
9/1/19	19130	33.99	33.99	33.99	35030	.10	.10	.10	Weaver	✓
9/8/19	19130	33.99	34.00	34.00	35030	.10	.09	.10	Weaver	

* - Please enter a ✓ to denote simulator solution change taking place