

Infrared Instrument
Weekly Calibration Verification Test Record

Department: MCSO

Instrument Serial #: 136506

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
11/25/18	18330	33.94	37.94	33.94	53030	.10	.10	.10	Weaver	
12/11/18	18310	33.91	34.00	34.00	53030	.10	.10	.10	Weaver	✓
12/19/18	18310	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
12/14/18	18310	34.00	34.00	34.00	53030	.09	.10	.10	Weaver	
12/27/18	18310	33.99	34.00	34.00	53030	.10	.10	.10	Weaver	
12/30/18	18310	33.99	33.99	34.00	53030	.09	.09	.09	Weaver	
1/3/19	18310	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	✓
1/10/19	18310	34.00	34.00	34.00	53030	.09	.10	.10	Weaver	
1/20/19	18310	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	
1/27/19	18310	34.00	34.00	34.00	53030	.10	.09	.10	Weaver	
2/1/19	18360	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	✓
2/10/19	18360	34.00	34.00	34.00	53030	.10	.10	.10	Weaver	

* - Please enter a ✓ to denote simulator solution change taking place