



MADISON COUNTY DISTRICT ATTORNEY'S OFFICE

CALIBRATION RECORDS SUBMISSION FORM FOR
C.P.L. ARTICLE 245 COMPLIANCE

AGENCY:

NYS University Police-Morrisville

CALIBRATION
RECORDS FOR
THE PERIOD:

January, 2019 to December, 2019

AGENCY CONTACT:

(If there are any questions
on the calibration records
submitted.)

Lt. Conor Duffy

Phone:

(315)684-6410

Email:

duffycj@morrisville.edu

PLEASE SUBMIT ALL CALIBRATION RECORDS BY A LABELED PDF SENT BY
EMAIL TO discovery@madisoncounty.ny.gov WITH THIS SUBMISSION FORM AS THE
FIRST PAGE OF THE PDF. THANK YOU!

PLEASE CHECK ONE:



With the present submission, the calibration records for the period specified above, that appear as a listed item of "automatic discovery" under C.P.L. § 245.20(1), are complete and have been submitted to the District Attorney's Office.



With the present submission, the calibration records for the period specified above, that are a listed item of "automatic discovery" found under C.P.L. § 245.20(1) are not complete and further calibration records for the above listed dates will be submitted to the District Attorney's Office as soon as they are available.

PLEASE NOTE that under C.P.L. § 245.20(1), the list of materials to be turned over for discovery is phrased as "including but not limited to" and all discovery materials must be turned over within fifteen (15) days of arraignment. Please note that under C.P.L. § 245.55(2), that a standing on-going request has been made to your agency from the district attorney by which you are required to turn over "a complete copy of its complete records and files related to the investigation of the case or the prosecution of the defendant".

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: NYS University Police

Instrument Serial #: 100123

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
12/21/18	18230	34.00	34.00	34.00	MP400	.10	.10	.10	Duffy	
12/28/18	18230	34.00	34.00	34.00	MP400	.10	.10	.10	MACKO	
1/4/19	18310	34.00	34.00	34.00	MP400	.10	.10	.10	Duffy	✓
1/14/19	18310	33.99	34.00	34.00	MP400	.09	.09	.10	Duffy	
1/18/19	18310	34.00	33.99	33.99	MP400	.10	.10	.10	Duffy	
1/25/19	18310	34.00	34.00	34.00	MP400	.10	.10	.10	Duffy	
2/1/19	18310	33.99	33.99	34.00	MP400	.09	.09	.10	Duffy	✓
2/8/19	18310	34.00	34.00	34.00	MP400	.10	.10	.10	MACKO	
2/15/19	18310	33.99	33.99	33.99	MP400	.10	.10	.10	Duffy	
2/22/19	18310	33.99	33.99	33.99	MP400	.10	.10	.10	MACKO	
3/1/19	18360	34.00	33.99	34.00	MP400	.10	.10	.10	Duffy	✓
3/14/19	18360	34.00	34.00	34.00	MP400	.10	.10	.10	Duffy	

* - Please enter a ✓ to denote simulator solution change taking place

END

2018

2019

BEGIN

* 3/9/19
No operator
available ML
and Training

SOLUTION CHANGE


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 01/04/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:48
INTERNAL STANDARD	VERIFIED	13:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:49
BLANK TEST	0.00	13:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:50
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:51
BLANK TEST	0.00	13:52



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 01/14/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:07
INTERNAL STANDARD	VERIFIED	14:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:08
BLANK TEST	0.00	14:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:09
BLANK TEST	0.00	14:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:11



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 01/18/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:43
INTERNAL STANDARD	VERIFIED	18:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:44
BLANK TEST	0.00	18:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:45
BLANK TEST	0.00	18:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:46
BLANK TEST	0.00	18:47



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 01/25/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:32
INTERNAL STANDARD	VERIFIED	00:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:33
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:34
BLANK TEST	0.00	00:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:35
BLANK TEST	0.00	00:36



SOLUTION CHANGE


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 02/01/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:36
INTERNAL STANDARD	VERIFIED	21:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:37
BLANK TEST	0.00	21:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:38
BLANK TEST	0.00	21:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:39
BLANK TEST	0.00	21:40



SUPERVISOR TEST



NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 02/08/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:15
INTERNAL STANDARD	VERIFIED	09:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:17
BLANK TEST	0.00	09:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 02/15/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:55
INTERNAL STANDARD	VERIFIED	16:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:56
BLANK TEST	0.00	16:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:57
BLANK TEST	0.00	16:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:59
BLANK TEST	0.00	16:59



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 02/22/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:19
INTERNAL STANDARD	VERIFIED	08:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23



SOLUTION CHANGE




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 03/01/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:56
INTERNAL STANDARD	VERIFIED	15:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:56
BLANK TEST	0.00	15:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:59
BLANK TEST	0.00	16:00



SUPERVISOR TEST



NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 03/16/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:59
INTERNAL STANDARD	VERIFIED	11:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:00
BLANK TEST	0.00	12:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:01
BLANK TEST	0.00	12:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:02
BLANK TEST	0.00	12:03



**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: NYS University Police

Instrument Serial #: 100123

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
3/22/19	18360	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
3/29/19	18360	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
4/5/19	18360	33.99	34.00	34.00	MP4400	.10	.10	.10	MACKO	✓
4/12/19	18360	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
4/19/19	18360	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
4/26/19	18360	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
5/3/19	19030	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	✓
5/10/19	19030	34.01	34.01	34.01	MP4400	.10	.10	.10	Duffy	
5/18/19	19030	34.00	34.00	34.00	MP4400	.10	.10	.10	Morgan	
5/24/19	19030	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
5/31/19	19030	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
6/7/19	19030	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	✓

* - Please enter a ✓ to denote simulator solution change taking place

SUPERVISOR TEST



 NYS UNIVERSITY POLICE
 SUNY MORRISVILLE
 DATAMASTER dmt: 100123

Date: 03/22/2019

OPERATOR NAME:
 JONATHAN S MACKO
 Permit #: 45982
 Operator Agency/Dept: SUC MORRISVILLE
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:26
INTERNAL STANDARD	VERIFIED	09:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:26
BLANK TEST	0.00	09:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:27
BLANK TEST	0.00	09:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:29
BLANK TEST	0.00	09:30



SUPERVISOR TEST




 NYS UNIVERSITY POLICE
 SUNY MORRISVILLE
 DATAMASTER dmt: 100123

Date: 03/29/2019

OPERATOR NAME:
 CONOR J DUFFY
 Permit #: 39998
 Operator Agency/Dept: SUC MORRISVILLE
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:50
INTERNAL STANDARD	VERIFIED	11:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:51
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:52
BLANK TEST	0.00	11:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:53
BLANK TEST	0.00	11:54



SOLUTION CHANGE




 NYS UNIVERSITY POLICE
 SUNY MORRISVILLE
 DATAMASTER dmt: 100123

Date: 04/05/2019

OPERATOR NAME:
 JONATHAN S MACKO
 Permit #: 45982
 Operator Agency/Dept: SUC MORRISVILLE
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:40
INTERNAL STANDARD	VERIFIED	08:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:41
BLANK TEST	0.00	08:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:42
BLANK TEST	0.00	08:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:45



SUPERVISOR TEST



 NYS UNIVERSITY POLICE
 SUNY MORRISVILLE
 DATAMASTER dmt: 100123

Date: 04/12/2019

OPERATOR NAME:
 CONOR J DUFFY
 Permit #: 39998
 Operator Agency/Dept: SUC MORRISVILLE
 Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:06
INTERNAL STANDARD	VERIFIED	18:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:06
BLANK TEST	0.00	18:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:07
BLANK TEST	0.00	18:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:09
BLANK TEST	0.00	18:10



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 04/19/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:05
INTERNAL STANDARD	VERIFIED	07:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:06
BLANK TEST	0.00	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 04/26/2019

OPERATOR NAME:
CONOR J. DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:16
INTERNAL STANDARD	VERIFIED	15:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:17
BLANK TEST	0.00	15:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:18
BLANK TEST	0.00	15:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:20
BLANK TEST	0.00	15:21



SOLUTION CHANGE



NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 05/03/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:13
INTERNAL STANDARD	VERIFIED	08:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 05/10/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:46
INTERNAL STANDARD	VERIFIED	22:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:47
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:49
BLANK TEST	0.00	22:50



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 05/18/2019

OPERATOR NAME:
TYLER J MORGAN
Permit #: 45936
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:03
INTERNAL STANDARD	VERIFIED	20:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:03
BLANK TEST	0.00	20:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 05/24/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:22
INTERNAL STANDARD	VERIFIED	07:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:24
BLANK TEST	0.00	07:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:25
BLANK TEST	0.00	07:26



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 05/31/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:37
INTERNAL STANDARD	VERIFIED	09:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:38
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:40
BLANK TEST	0.00	09:41



SOLUTION CHANGE




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 06/07/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:12
INTERNAL STANDARD	VERIFIED	19:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:12
BLANK TEST	0.00	19:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:14
BLANK TEST	0.00	19:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:15
BLANK TEST	0.00	19:16



**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: NYS University Police Instrument Serial #: 100123

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
6/11/19	19090	34.00	34.00	34.00	MP400	0.20	0.10	0.10	Duffy	✓
6/14/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
6/26/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
6/28/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
7/10/19	19090	34.0	34.0	34.0	MP4400	.10	.10	.10	Wright	✓
7/12/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
7/19/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
7/20/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
8/2/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	✓
8/9/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	Morgan	
8/16/19	19090	34.00	34.00	34.00	MP4400	.10	.10	.10	Duffy	
8/28/19	19090	34.0	34.0	34.0	MP4400	.10	.10	.10	Wright	

* - Please enter a ✓ to denote simulator solution change taking place

Solution from
DCJS for
Annual
Calibration
f. No operator
avail on 6/21/19

* No operator
avail on 7/5/19

SOLUTION CHANGE




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 06/11/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:21
INTERNAL STANDARD	VERIFIED	16:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:21
BLANK TEST	0.00	16:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:23
BLANK TEST	0.00	16:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:24
BLANK TEST	0.00	16:25



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 06/14/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:50
INTERNAL STANDARD	VERIFIED	16:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:51
BLANK TEST	0.00	16:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:52
BLANK TEST	0.00	16:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:53
BLANK TEST	0.00	16:54



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 06/26/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:29
INTERNAL STANDARD	VERIFIED	11:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:30
BLANK TEST	0.00	11:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:31
BLANK TEST	0.00	11:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:32
BLANK TEST	0.00	11:33



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 06/28/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:45
INTERNAL STANDARD	VERIFIED	18:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:46
BLANK TEST	0.00	18:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:47
BLANK TEST	0.00	18:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:48
BLANK TEST	0.00	18:49



SOLUTION CHANGE




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 07/10/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:10
INTERNAL STANDARD	VERIFIED	11:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:10
BLANK TEST	0.00	11:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:11
BLANK TEST	0.00	11:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:13
BLANK TEST	0.00	11:13



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 07/12/2019

OPERATOR NAME:
JONATAHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:14
INTERNAL STANDARD	VERIFIED	14:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:17
BLANK TEST	0.00	14:18



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 07/19/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:43
INTERNAL STANDARD	VERIFIED	11:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:44
BLANK TEST	0.00	11:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:45
BLANK TEST	0.00	11:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:46
BLANK TEST	0.00	11:47



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 07/26/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:39
INTERNAL STANDARD	VERIFIED	17:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:39
BLANK TEST	0.00	17:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:41
BLANK TEST	0.00	17:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43



SOLUTION CHANGE



NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 08/02/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:12
INTERNAL STANDARD	VERIFIED	08:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 08/09/2019

OPERATOR NAME:
TYLER J MORGAN
Permit #: 45936
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:25
INTERNAL STANDARD	VERIFIED	17:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:26
BLANK TEST	0.00	17:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:27
BLANK TEST	0.00	17:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:28
BLANK TEST	0.00	17:29



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 08/16/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:19
INTERNAL STANDARD	VERIFIED	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:22



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 08/23/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 31034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:41
INTERNAL STANDARD	VERIFIED	07:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45



**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: NYSUP Morrisville

Instrument Serial #: 100123

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
8/30/19	19096	34.0	34.0	34.0	MP4400 100123	.10	.10	.10	Wright	
9/7/19	19090	34.0	34.0	34.0	MP4400 100123	.10	.10	.10	Morgan	
9/16/19	19130	34.0	34.0	34.0	MP4400	.10	.10	.10	Duffy	X
9/13	19130	34.00	34.00	34.00	MP4400	.10	.10	.10	Wright	
9/20	19130	34.0	34.0	34.0	MP4400	.10	.10	.10	Wright	
9/27	19130	34.0	34.0	34.0	MP4400	.10	.10	.10	Wright	
10/4	19180	34.0	34.0	34.0	MP4400	.10	.10	.10	Morgan	X
10/14	19180	34.00	34.00	34.00	MP4400	.10	.10	.10	Morgan	
10/20	19180	34.00	34.00	34.00	MP4400	.10	.10	.10	Morgan	
10/25	19180	34.0	34.0	34.0	MP4400	.10	.10	.10	Wright	
11/1	19180	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	✓
11/18	19180	33.99	34.00	34.00	MP4400	.10	.10	.10	Netzbond	

* - Please enter a ✓ to denote simulator solution change taking place

* Performed by mistake*

Solution change

SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 08/30/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:46
INTERNAL STANDARD	VERIFIED	14:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:47
BLANK TEST	0.00	14:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:48
BLANK TEST	0.00	14:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:49
BLANK TEST	0.00	14:50



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 09/07/2019

OPERATOR NAME:
TYLER J MORGAN
Permit #: 45936
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:51
INTERNAL STANDARD	VERIFIED	18:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:52
BLANK TEST	0.00	18:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:53
BLANK TEST	0.00	18:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:54
BLANK TEST	0.00	18:55



SOLUTION CHANGE


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 09/08/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:33
INTERNAL STANDARD	VERIFIED	07:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:36
BLANK TEST	0.00	07:36



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 09/13/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:37
INTERNAL STANDARD	VERIFIED	18:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:37
BLANK TEST	0.00	18:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:38
BLANK TEST	0.00	18:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:40
BLANK TEST	0.00	18:40



SUPERVISOR TEST



NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 09/20/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:50
INTERNAL STANDARD	VERIFIED	14:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:50
BLANK TEST	0.00	14:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:51
BLANK TEST	0.00	14:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:53
BLANK TEST	0.00	14:53



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 09/27/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:33
INTERNAL STANDARD	VERIFIED	15:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:33
BLANK TEST	0.00	15:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:35
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:36
BLANK TEST	0.00	15:37



SOLUTION CHANGE




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 10/04/2019

OPERATOR NAME:
TYLER J MORGAN
Permit #: 45936
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:56
INTERNAL STANDARD	VERIFIED	22:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:57
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:59
BLANK TEST	0.00	23:00



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 10/11/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:55
INTERNAL STANDARD	VERIFIED	14:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:55
BLANK TEST	0.00	14:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:57
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:58
BLANK TEST	0.00	14:59



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 10/14/2019

OPERATOR NAME:
TYLER J MORGAN
Permit #: 45936
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:24
INTERNAL STANDARD	VERIFIED	19:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:25
BLANK TEST	0.00	19:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:26
BLANK TEST	0.00	19:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:27
BLANK TEST	0.00	19:28



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 10/20/2019

OPERATOR NAME:
TYLER J MORGAN
Permit #: 45936
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:38
INTERNAL STANDARD	VERIFIED	00:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:39
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:41
BLANK TEST	0.00	00:42



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 10/25/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:14
INTERNAL STANDARD	VERIFIED	15:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:15
BLANK TEST	0.00	15:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:16
BLANK TEST	0.00	15:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:17
BLANK TEST	0.00	15:18



SOLUTION CHANGE


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 11/01/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:56
INTERNAL STANDARD	VERIFIED	12:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:56
BLANK TEST	0.00	12:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:57
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:59
BLANK TEST	0.00	12:59



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 11/08/2019

OPERATOR NAME:
PHILIP B NETZBAND
Permit #: 40415
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:40
INTERNAL STANDARD	VERIFIED	20:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:40
BLANK TEST	0.00	20:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:41
BLANK TEST	0.00	20:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:43
BLANK TEST	0.00	20:43



**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: NYSAP Morrisville

Instrument Serial #: 100123

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
11/15/19	19180	34.00	33.99	34.00	MP4400	.10	.10	.10	MACKO	
11/23/19	19180	33.99	34.00	34.00	MP4400	.10	.10	.10	Netzeland	
11/29/19	19180	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
12/3/19	19250	33.99	33.99	34.00	MP4400	.10	.10	.10	DUFFY	X
12/6/19	19250	34.00	34.00	34.00	MP4400	.10	.10	.10	Wright	
12/13/19	19250	34.00	34.00	33.99	MP4400	.10	.10	.10	MACKO	
12/21/19	19250	34.00	34.00	34.00	MP4400	.10	.10	.10	Wright	
12/27/19	19250	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	
1/8/20	19250	34.00	34.00	34.00	MP4400	.10	.10	.10	MACKO	X
1/14/20										
1/20/20										

Solution change

END
2019

* - Please enter a √ to denote simulator solution change taking place

SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 11/15/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:15
INTERNAL STANDARD	VERIFIED	08:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:18
BLANK TEST	0.00	08:19



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 11/23/2019

OPERATOR NAME:
PHILIP B NETZBAND
Permit #: 40415
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:06
INTERNAL STANDARD	VERIFIED	02:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:07
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:09
BLANK TEST	0.00	02:10



SUPERVISOR TEST


NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 11/29/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:33
INTERNAL STANDARD	VERIFIED	14:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:34
BLANK TEST	0.00	14:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:35
BLANK TEST	0.00	14:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:36
BLANK TEST	0.00	14:37



SOLUTION CHANGE



NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 12/03/2019

OPERATOR NAME:
CONOR J DUFFY
Permit #: 39998
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:47
INTERNAL STANDARD	VERIFIED	07:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:47
BLANK TEST	0.00	07:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:48
BLANK TEST	0.00	07:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:50
BLANK TEST	0.00	07:50



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 12/06/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:14
INTERNAL STANDARD	VERIFIED	16:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:15
BLANK TEST	0.00	16:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:16
BLANK TEST	0.00	16:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:17
BLANK TEST	0.00	16:18



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 12/13/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:11
INTERNAL STANDARD	VERIFIED	09:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:11
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:12
BLANK TEST	0.00	09:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:14
BLANK TEST	0.00	09:14



SUPERVISOR TEST




NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 12/21/2019

OPERATOR NAME:
NICOLE M WRIGHT
Permit #: 41034
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:31
INTERNAL STANDARD	VERIFIED	11:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:32
BLANK TEST	0.00	11:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:33
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:34
BLANK TEST	0.00	11:35



SUPERVISOR TEST

NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 12/27/2019

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:04
INTERNAL STANDARD	VERIFIED	17:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:05
BLANK TEST	0.00	17:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:07
BLANK TEST	0.00	17:08

