

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
10/4/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
10/12/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
10/27/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
10/29/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
10/29/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
11/05/18	18230	33.99c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	SOLUTION CHANGE BOT#900	1/26/19
11/09/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
11/16/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
11/28/18	18230	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	1/26/19
12/3/18	18310	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	SOLUTION CHANGE BOT#492	3/24/19
12/9/18	18310	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	2/24/19
12/15/18	18310	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	3/24/19
12/29/18	18310	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	WEEKLY SOLUTION	3/24/19

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
7/24/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	WEEKLY SOLUTION	9/26/18
7/30/18	18100	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	Lowell	WEEKLY SOLUTION	9/26/18
8/1/18	18170	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	NEW LOT 18170 BOT # 155	11/30/18
8/5/18	18170	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	WEEKLY SOLUTION	11/30/18
8/12/18	18170	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	WEEKLY SOLUTION	11/30/18
8/18/18	18170	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	WEEKLY SOLUTION	11/30/18
8/23/18	18170	33.99 ^c	.09	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	WEEKLY SOLUTION	11/30/18
8/29/18	18170	33.99 ^c	.10	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	Lowell	WEEKLY SOLUTION	11/30/18
9/2/18	18170	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	SOLUTION 18170 CHANGE BOT # 104	11/30/18
9/16/18	18170	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	Lowell	WEEKLY SOLUTION	11/30/18
9/22/18	18170	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.09	Lowell	WEEKLY SOLUTION	11/30/18
9/24/18	* * *	* * *	OT FOR CALIBRATION	FOR CALIBRATION	TO	ALBANY *	* * *	* * *	* * *	ALBANY *	* * *	* * *
10/01/18	18230	33.99 ^c	.09	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.09	Lowell	BACK IN SUES NEW LOT BOT # 883	1/26/19

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
5/2/18	18040	33.99 ^c	.10	34.00 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION CHANGE WEEKLY	BOF # 185	7/29/18	
5/8/18	18040	33.99 ^c	.10	33.99 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION WEEKLY		7/29/18	
5/15/18	18040	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		7/29/18	
5/22/18	18040	33.99 ^c	.10	33.99 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION WEEKLY		7/29/18	
5/27/18	18040	33.99 ^c	.09	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		7/29/18	
6/3/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION CHANGE BOF #377		9/26/18	
6/9/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		9/26/18	
6/13/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		9/26/18	
6/18/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		9/26/18	
6/24/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		9/26/18	
7/1/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		9/26/18	
7/1/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION CHANGE BOF #377		9/26/18	
7/9/18	18100	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION WEEKLY		9/26/18	

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
1/29/18	17310	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION	3/25/18
2/4/18	17310	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION	3/25/18
2/15/18	17310	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION	3/25/18
2/27/18	17310	34.00 ^c	.09	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.09	LOWELL	SOLUTION	3/25/18
3/5/18	17310	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION CHANGE	3/25/18
3/17/18	17310	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION	3/25/18
3/27/18	17310	34.00 ^c	.10	34.00 ^c	.09	34.00 ^c	.10	34.00 ^c	.10	LOWELL	SOLUTION	3/25/18
4/3/18	18040	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	NEW SOLUTION BOT.# LOT CHANGE 21	7/29/18
4/16/18	18040	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	WEEKLY	7/29/18
4/11/18	18040	34.00 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	WEEKLY	7/29/18
4/16/18	18040	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	33.99 ^c	.10	LOWELL	WEEKLY	7/29/18
4/22/18	18040	33.99 ^c	.10	33.99 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	WEEKLY	7/29/18
4/24/18	18040	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	34.00 ^c	.10	LOWELL	WEEKLY	7/29/18

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
12/12/16	16240	34.02°	.09	34.01°	.09	34.00°	.09	34.00°	.09	THOMPSON	SOLUTION	03/21/17
12/16/16	16240	34.00°	.09	34.01°	.09	34.01°	.09	34.01°	.09	THOMPSON	SOLUTION	03/21/17
12/30/16	16240	34.02°	.09	34.00°	.09	34.01°	.09	34.01°	.09	THOMPSON	SOLUTION	03/21/17
01/02/17	16240	34.01°	.09	34.01°	.09	33.98°	.09	33.98°	.09	THOMPSON	NEW SOLIT	03/21/17
01/02/17	16240	34.01°	.09	34.02°	.09	34.02°	.09	34.02°	.09	THOMPSON	SOLUTION	03/21/17
01/06/17	16240	34.01°	.09	34.02°	.09	34.00°	.09	34.00°	.09	THOMPSON	SOLUTION	03/21/17
01/14/17	16240	34.01°	.09	34.01°	.09	34.01°	.09	34.01°	.09	THOMPSON	SOLUTION	03/21/17
01/23/17	16240	34.01°	.09	34.02°	.09	34.01°	.09	34.01°	.09	THOMPSON	SOLUTION	03/21/17
01/28/17	16240	34.01°	.09	34.03°	.09	34.02°	.09	34.02°	.09	THOMPSON	SOLUTION	03/21/17
02/01/17	16380	34.00°	.09	34.01°	.09	34.02°	.09	34.02°	.09	THOMPSON	NEW LOT	05/28/17
02/07/17	16380	34.01°	.09	34.01°	.09	34.00°	.09	34.00°	.09	THOMPSON	SOLUTION	05/28/17
1/20/18	17310	34.00°	.10	34.00°	.10	34.00°	.10	34.00°	.10	LOWELL	NEW LOT	3/25/18
1/18/18	—	—	UNIT	TO	DJSS	FOR	CALIBRATION	—	—	—	SAME DAY BACK IN SERVICE	—

LATIC ENTRY