

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
10/17/19	19180	34.00C	.10	34.00C	.10	34.00C	.10	34.00C	.10	LOWELL	WEEKLY SOLUTION	1/24/20
10/23/19	19180	33.99C	.09	33.99C	.09	34.00C	.09	34.00C	.09	LOWELL	WEEKLY SOLUTION	1/24/20
10/29/19	19180	33.99C	.09	33.99C	.09	33.99C	.09	33.99C	.09	LOWELL	WEEKLY SOLUTION	1/24/20
11/1/19	19180	34.00C	.10	34.00C	.10	34.00C	.10	34.00C	.10	LOWELL	WEEKLY SOLUTION	1/24/20
11/5/19	19180	34.00C	.10	34.00C	.10	34.00C	.10	34.00C	.10	LOWELL	WEEKLY SOLUTION	1/24/20
11/6/19	19180	34.00C	.10	34.00C	.10	34.00C	.10	34.00C	.10	LOWELL	WEEKLY SOLUTION	1/24/20
11/11/19	19180	34.00C	.10	34.00C	.10	34.00C	.10	34.00C	.10	LOWELL	WEEKLY SOLUTION	1/24/20
11/17/19	19180	34.00C	.09	33.99C	.09	34.00C	.09	54.00C	.09	LOWELL	WEEKLY SOLUTION	1/24/20
11/22/19	19180	34.00C	.09	34.00C	.09	34.00C	.09	34.00C	.09	LOWELL	WEEKLY SOLUTION	1/24/20
12/2/19	19180	34.00C	0.09	34.00C	0.09	34.00C	0.09	34.00C	0.09	BAKST	WEEKLY SOLUTION	1/24/20
12/3/19	19250	34.00C	.09	34.00C	.10	34.00C	.09	34.00C	.09	LOWELL	SOLUTION CHANGE, LOT, BATCH 3/30/2020	3/30/20
12/7/19	19250	34.00C	.10	34.00C	.09	34.00C	.10	34.00C	.09	LOWELL	WEEKLY SOLUTION	3/30/20
12/10/19	19250	34.00C	.09	34.00C	.10	34.00C	.09	34.00C	.10	LOWELL	WEEKLY SOLUTION	3/30/20

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
8/25/19	19130	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	11/28/19
8/30/19	19130	33.99 ^c	.07	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	11/28/19
9/1/19	19130	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	SOLUTION CHANGE BOT 87A	11/28/19
9/5/19	19130	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	11/28/19
9/11/19	19130	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	11/28/19
9/17/19	19130	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	11/28/19
9/24/19	19130	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	11/28/19
9/25/19	** *	DATAMASTER	OUT	DATAMASTER	TO	ALBANY	FOR	ALBANY	FOR	CALIBRATION	** *	** *
9/25/19	19180	33.99 ^c	.10	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	IN SVC. FROM ALBANY. BOT 740	1/24/20
9/29/19	19180	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	1/24/20
10/5/19	19180	33.99 ^c	.10	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	SOLUTION CHANGE BOT 130	1/24/20
10/5/19	19180	33.99 ^c	.09	33.99 ^c	.10	33.99 ^c	.10	33.99 ^c	.10	LOWELL	SOLUTION TEST CHANGE BOT 130	1/24/20
10/11/19	19180	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	33.99 ^c	.09	LOWELL	WEEKLY SOLUTION	1/24/20

RFI

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
6/10/19	19090	34.00c	.09	34.00c	.09	34.00c	.09	Lowell	Weekly SOLUTION	9/20/19
6/13/19	19090	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	9/20/19
6/17/19	19090	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	9/20/19
6/25/19	19090	33.99c	.09	33.99c	.05	33.99c	.09	Lowell	Weekly SOLUTION	9/20/19
7/1/19	19090	33.99	.09	33.99	.09	33.99	.09	Barres	Weekly SOLUTION	9/20/19
7/13/19	19090	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	9/20/19
7/20/19	19090	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	9/20/19
7/25/19	19090	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	9/20/19
8/1/19	19130	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	SOLUTION BOT #765 CHANGE LOT	11/28/19
8/3/19	19130	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	11/28/19
8/6/19	19130	33.99c	.07	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	11/28/19
8/12/19	19130	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	11/28/19
8/18/19	19130	33.99c	.09	33.99c	.09	33.99c	.09	Lowell	Weekly SOLUTION	11/24/19

Sol. change
Out 475

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
3/22/19	18360	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	5/28/19
3/27/19	18360	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	5/28/19
4/2/19	19030	33.99c	.09	33.99c	.09	34.00c	.09	34.00c	.09	LOWELL	SOLUTION CHANGE, BOT 730	7/21/19
4/9/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
4/15/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
4/21/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
4/27/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
5/3/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	SOLUTION CHANGE BOT 745	7/21/19
5/8/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
5/15/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
5/21/19	19030	33.99c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
5/27/19	19030	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	7/21/19
6/1/19	19090	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	SOLUTION CHANGE, BOT 43	9/20/19

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
1/2/19	18310	34.00c	.09	34.00c	.09	34.00	.09	LOWELL	SOLUTION CHANGE BOT 568	LOWELL	SOLUTION CHANGE BOT 568	3/24/19
1/8/19	18310	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	3/24/19
1/14/19	18310	33.99c	.09	33.99c	.09	33.99c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	3/24/19
1/20/19	18310	33.99c	.09	33.99c	.09	33.99c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	3/24/19
1/26/19	18310	33.99c	.09	33.99c	.09	33.99c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	3/24/19
2/1/19	18360	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	SOLUTION CHANGE BOT 307	LOWELL	SOLUTION CHANGE BOT 307	5/28/19
2/7/19	18360	34.00c	.10	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	5/28/19
2/13/19	18360	33.99c	.09	33.99c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	5/28/19
2/19/19	18360	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	5/28/19
2/26/19	18360	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	5/28/19
3/4/19	18360	33.99c	.09	33.99c	.09	33.99c	.09	LOWELL	SOLUTION CHANGE BOT 307	LOWELL	SOLUTION CHANGE BOT 307	5/28/19
3/9/19	18360	33.99c	.09	33.99c	.09	33.99c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	5/28/19
3/16/19	18360	33.99c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	LOWELL	WEEKLY SOLUTION	5/28/19