

SOLUTION CHANGE

-----  
 CANASTOTA VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 138406  
 -----

Date: 02/03/2020

OPERATOR NAME:  
 WILLIAM E PREUSS  
 Permit #: 39320  
 Operator Agency/Dept: CANASTOTA VILL. PD  
 Reference Standard #: 19350

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 23:39 |
| INTERNAL STANDARD       | VERIFIED | 23:39 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 23:40 |
| BLANK TEST              | 0.00     | 23:41 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 23:41 |
| BLANK TEST              | 0.00     | 23:42 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 23:42 |
| BLANK TEST              | 0.00     | 23:43 |

