

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Coastal Police Dept

Instrument Serial #: 138406

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
1/02/2020	19250	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	✓
1/05/2020	19250	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	

* - Please enter a ✓ to denote simulator solution change taking place

SOLUTION CHANGE

CANASTOTA VILLAGE
POLICE DEPT
DATAMASTER dmt: 138406

Date: 01/02/2020

OPERATOR NAME:
WILLIAM E PREUSS
Permit #: 39320
Operator Agency/Dept: CANASTOTA VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:51
INTERNAL STANDARD	VERIFIED	22:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:51
BLANK TEST	0.00	22:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:53
BLANK TEST	0.00	22:54

