

Infrared Instrument  
Weekly Calibration Verification Test Record

Department: Coastal Police Dept

Instrument Serial #: 138406

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
1/02/2020	19250	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	✓
1/05/2020	19250	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	

\* - Please enter a ✓ to denote simulator solution change taking place

SUPERVISOR TEST

CANASTOTA VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138406

Date: 01/05/2020

OPERATOR NAME:  
WILLIAM E PREUSS  
Permit #: 39320  
Operator Agency/Dept: CANASTOTA VILL. PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:37
INTERNAL STANDARD	VERIFIED	22:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:37
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40

