

**Preliminary Breath Test Device
Monthly Verification Log**

Department Catskill Police Dept

Device Serial # 138406

Test Date	Sim. Solution Lot #	Simulator Temperature			Sim. Serial #	Test Results			Officer Performing Verification	Sim. Solution Change *
		1	2	3		1	2	3		

5/19/19	19030	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
5/27/19	19030	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
6/3/19	19090	34.02	34.02	34.02	MP4042	.09	.09	.10	Preuss	✓
6/10/19	19090	34.02	34.02	34.02	MP4042	.09	.09	.10	Preuss	
6/18/19	19090	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
6/23/19	19090	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
7/1/19	19090	34.02	34.02	34.02	MP4042	.09	.10	.10	Preuss	✓
7/9/19	19090	Chopper Frequency Low			MP4042	Chopper Frequency Low			Preuss	
7/14/19	19090	Chopper Frequency Low			MP4042	Chopper Frequency Low			Preuss	
7/28/19	19090	34.01	34.01	34.01	MP4042	.09	.10	.09	Preuss	
8/2/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	✓
8/4/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
8/11/19	19130	34.02	34.02	34.01	MP4042	.09	.09	.09	Preuss	
8/18/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
8/25/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	

* - Please enter a ✓ to denote a simulator solution lot number change

Version 8/12/2015

SUPERVISOR TEST

CANASTOTA VILLAGE
POLICE DEPT
DATAMASTER dmt: 138406

Date: 08/04/2019

OPERATOR NAME:
WILLIAM E PREUSS
Permit #: 39320
Operator Agency/Dept: CANASTOTA VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:29
INTERNAL STANDARD	VERIFIED	22:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:30
BLANK TEST	0.00	22:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:31
BLANK TEST	0.00	22:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:32
BLANK TEST	0.00	22:33

