

**Preliminary Breath Test Device
Monthly Verification Log**

Department Catskill Police Dept

Device Serial # 138406

| Test Date | Sim. Solution Lot # | Simulator Temperature | | | Sim. Serial # | Test Results | | | Officer Performing Verification | Sim. Solution Change * |
|-----------|---------------------|-----------------------|---|---|---------------|--------------|---|---|---------------------------------|------------------------|
| | | 1 | 2 | 3 | | 1 | 2 | 3 | | |

| | | | | | | | | | | |
|---------|-------|-----------------------|-------|-------|--------|-----------------------|-----|-----|--------|---|
| 5/19/19 | 19030 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 5/27/19 | 19030 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 6/3/19 | 19090 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .10 | Preuss | ✓ |
| 6/10/19 | 19090 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .10 | Preuss | |
| 6/18/19 | 19090 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 6/23/19 | 19090 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 7/1/19 | 19090 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .10 | .10 | Preuss | ✓ |
| 7/9/19 | 19090 | Chopper Frequency Low | | | MP4042 | Chopper Frequency Low | | | Preuss | |
| 7/14/19 | 19090 | Chopper Frequency Low | | | MP4042 | Chopper Frequency Low | | | Preuss | |
| 7/28/19 | 19090 | 34.01 | 34.01 | 34.01 | MP4042 | .09 | .10 | .09 | Preuss | |
| 8/2/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | ✓ |
| 8/4/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 8/11/19 | 19130 | 34.02 | 34.02 | 34.01 | MP4042 | .09 | .09 | .09 | Preuss | |
| 8/19/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 8/25/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |

* - Please enter a ✓ to denote a simulator solution lot number change

Version 8/12/2015

SUPERVISOR TEST

CANASTOTA VILLAGE
POLICE DEPT
DATAMASTER dmt: 138406

Date: 08/11/2019

OPERATOR NAME:
WILLIAM E PREUSS
Permit #: 39320
Operator Agency/Dept: CANASTOTA VILL. PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 22:45 |
| INTERNAL STANDARD | VERIFIED | 22:45 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 22:45 |
| BLANK TEST | 0.00 | 22:46 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 22:46 |
| BLANK TEST | 0.00 | 22:47 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 22:48 |
| BLANK TEST | 0.00 | 22:48 |

