

Infrared Instrument  
Weekly Calibration Verification Test Record

Department: Connetquot Police Dept

Instrument Serial #: 138406

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
9/2/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	✓
9/8/19	19130	34.02	34.02	34.02	MP4042	.10	.09	.09	Preuss	
9/15/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
9/22/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
10/01/19	19180	34.01	34.02	34.02	MP4042	.09	.09	.09	Preuss	✓
10/6/19	19180	34.02	34.01	34.02	MP4042	.09	.10	.09	Preuss	
10/13/19	19180	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
10/20/19	19180	34.02	34.01	34.02	MP4042	.09	.09	.09	Preuss	
10/28/19	19180	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
11/3/19	19180	34.02	34.02	34.02	MP4042	.09	.09	.10	Preuss	✓
11/10/19	19180	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	
11/17/19	19180	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	

\* - Please enter a ✓ to denote simulator solution change taking place

SUPERVISOR TEST

-----  
CANASTOTA VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138406  
-----

Date: 09/22/2019

OPERATOR NAME:  
WILLIAM E PREUSS  
Permit #: 39320  
Operator Agency/Dept: CANASTOTA VILL. PD  
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:21
INTERNAL STANDARD	VERIFIED	23:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:21
BLANK TEST	0.00	23:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:22
BLANK TEST	0.00	23:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:23
BLANK TEST	0.00	23:24

