

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Connetquot Police Dept

Instrument Serial #: 138406

| Test Date | Simulator Solution Lot # | Simulator Temperature | | | Sim. Serial# | Test Results | | | Breath Analysis Supervisor | Sim. Sol. Change * |
|-----------|--------------------------|-----------------------|-------|-------|--------------|--------------|-----|-----|----------------------------|--------------------|
| | | 1 | 2 | 3 | | 1 | 2 | 3 | | |
| 9/2/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | ✓ |
| 9/8/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .10 | .09 | .09 | Preuss | |
| 9/15/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 9/22/19 | 19130 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 10/01/19 | 19180 | 34.01 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | ✓ |
| 10/6/19 | 19180 | 34.02 | 34.01 | 34.02 | MP4042 | .09 | .10 | .09 | Preuss | |
| 10/13/19 | 19180 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 10/20/19 | 19180 | 34.02 | 34.01 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 10/28/19 | 19180 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .09 | Preuss | |
| 11/3/19 | 19180 | 34.02 | 34.02 | 34.02 | MP4042 | .09 | .09 | .10 | Preuss | ✓ |
| 11/10/19 | 19180 | 34.01 | 34.01 | 34.01 | MP4042 | .09 | .09 | .09 | Preuss | |
| 11/17/19 | 19180 | 34.01 | 34.01 | 34.01 | MP4042 | .09 | .09 | .09 | Preuss | |

* - Please enter a ✓ to denote simulator solution change taking place

SUPERVISOR TEST

CANASTOTA VILLAGE
POLICE DEPT
ATAMASTER dmt: 138406

Date: 10/20/2019

OPERATOR NAME:
WILLIAM E PREUSS
Permit #: 39320
Operator Agency/Dept: CANASTOTA VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 22:37 |
| INTERNAL STANDARD | VERIFIED | 22:37 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 22:37 |
| BLANK TEST | 0.00 | 22:38 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 22:38 |
| BLANK TEST | 0.00 | 22:39 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 22:39 |
| BLANK TEST | 0.00 | 22:40 |

