

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Connetquot Police Dept

Instrument Serial #: 138406

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
9/2/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	✓
9/8/19	19130	34.02	34.02	34.02	MP4042	.10	.09	.09	Preuss	
9/15/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
9/22/19	19130	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
10/01/19	19180	34.01	34.02	34.02	MP4042	.09	.09	.09	Preuss	✓
10/6/19	19180	34.02	34.01	34.02	MP4042	.09	.10	.09	Preuss	
10/13/19	19180	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
10/20/19	19180	34.02	34.01	34.02	MP4042	.09	.09	.09	Preuss	
10/28/19	19180	34.02	34.02	34.02	MP4042	.09	.09	.09	Preuss	
11/3/19	19180	34.02	34.02	34.02	MP4042	.09	.09	.10	Preuss	✓
11/10/19	19180	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	
11/17/19	19180	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	

* - Please enter a ✓ to denote simulator solution change taking place

SUPERVISOR TEST

CANASTOTA VILLAGE
POLICE DEPT
DATAMASTER dmt: 138406

Date: 10/27/2019

OPERATOR NAME:
WILLIAM E PREUSS
Permit #: 39320
Operator Agency/Dept: CANASTOTA VILL. PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:29
INTERNAL STANDARD	VERIFIED	22:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:29
BLANK TEST	0.00	22:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:30
BLANK TEST	0.00	22:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:31
BLANK TEST	0.00	22:32

