

Infrared Instrument
Weekly Calibration Verification Test Record

Department: Concordia Police Dept.

Instrument Serial #: 138406

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
11/24/19	19180	34.01	34.01	34.02	MP4042	.09	.09	.09	Preuss	
12/3/19	19250	34.01	34.01	34.01	MP4042	.09	.10	.09	Preuss	✓
12/8/19	19250	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	
12/15/19	19250	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	
12/22/19	19250	34.01	34.01	34.01	MP4042	.09	.09	.09	Preuss	

* - Please enter a ✓ to denote simulator solution change taking place

SUPERVISOR TEST

CANASTOTA VILLAGE
POLICE DEPT
DATAMASTER dmt: 138406

Date: 12/15/2019

OPERATOR NAME:
WILLIAM E PREUSS
Permit #: 39320
Operator Agency/Dept: CANASTOTA VILL. PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:33
INTERNAL STANDARD	VERIFIED	22:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:35
BLANK TEST	0.00	22:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:36
BLANK TEST	0.00	22:37

