



MADISON COUNTY DISTRICT ATTORNEY'S OFFICE

CALIBRATION RECORDS SUBMISSION FORM FOR C.P.L. ARTICLE 245 COMPLIANCE

AGENCY:

NYS University Police-Morrisville

CALIBRATION
RECORDS FOR
THE PERIOD:

January 20 to December 20

AGENCY CONTACT:

(If there are any questions
on the calibration records
submitted.)

Lt. Conor Duffy

Phone: (315)684-6410

Email: duffycj@morrisville.edu

**PLEASE SUBMIT ALL CALIBRATION RECORDS BY A LABELED PDF SENT BY
EMAIL TO discovery@madisoncounty.ny.gov WITH THIS SUBMISSION FORM AS THE
FIRST PAGE OF THE PDF. THANK YOU!**

PLEASE CHECK ONE:

With the present submission, the calibration records for the period specified above, that appear as a listed item of "automatic discovery" under C.P.L. § 245.20(1), are complete and have been submitted to the District Attorney's Office.

With the present submission, the calibration records for the period specified above, that are a listed item of "automatic discovery" found under C.P.L. § 245.20(1) are not complete and further calibration records for the above listed dates will be submitted to the District Attorney's Office as soon as they are available.

PLEASE NOTE that under C.P.L. § 245.20(1), the list of materials to be turned over for discovery is phrased as "including but not limited to" and all discovery materials must be turned over within fifteen (15) days of arraignment. Please note that under C.P.L. § 245.55(2), that a standing on-going request has been made to your agency from the district attorney by which you are required to turn over "a complete copy of its complete records and files related to the investigation of the case or the prosecution of the defendant".

Infrared Instrument
Weekly Calibration Verification Test Record

Department: NYSUP Morrisville

Instrument Serial #: 100123

* - Please enter a \checkmark to denote simulator solution change taking place

SOLUTION CHANGE

NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 01/06/2020

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:06
INTERNAL STANDARD	VERIFIED	08:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:08
BLANK TEST	0.00	08:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:09
BLANK TEST	0.00	08:10

