

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: NYSUP Morrisville

Instrument Serial #: 100123

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
1/6/20	19250	34.00	34.00	34.00	MP4460	.10	.10	.10	MACKO	X
1/10/20	19250	34.00	34.00	33.99	MP4400	.10	.10	.10	MACKO	

* - Please enter a √ to denote simulator solution change taking place

SUPERVISOR TEST

NYS UNIVERSITY POLICE
SUNY MORRISVILLE
DATAMASTER dmt: 100123

Date: 01/10/2020

OPERATOR NAME:
JONATHAN S MACKO
Permit #: 45982
Operator Agency/Dept: SUC MORRISVILLE
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:50
INTERNAL STANDARD	VERIFIED	11:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:50
BLANK TEST	0.00	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:51
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:53
BLANK TEST	0.00	11:54

