

SUPERVISOR TEST

ONEIDA CITY
POLICE DEPT
DATAMASTER dmt: 137806

Date: 02/02/2020

OPERATOR NAME:
STEVEN J LOWELL
Permit #: 40046
Operator Agency/Dept: ONEIDA CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:35
INTERNAL STANDARD	VERIFIED	23:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:36
BLANK TEST	0.00	23:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:37
BLANK TEST	0.00	23:38

