

## CITY OF ONEIDA POLICE DEPARTMENT CALIBRATION TEST RECORD

NATIONAL PATENT ANALYTICAL SYSTEMS INC.  
DATAMASTER DMT

SERIAL NUMBER 137806

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
12/16/19	19250	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	LOWELL	WEEKLY SOLUTION	3/30/20
12/22/19	19250	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	Mrsack	WEEKLY SOLUTION	3/30/20
12/27/19	19250	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	Mrsack	WEEKLY SOLUTION	3/30/20
12/31/19	19250	38.99 <sup>c</sup>	.10	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.10	LOWELL	SOLUTION CHANGE BOT #923	3/30/20
01/03/20	19250	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	LOWELL	WEEKLY SOLUTION	3/30/20
01/09/20	19250	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/15/20	19250	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/21/20	19250	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/27/20	19250	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/31/20	19350	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.10	LOWELL	SOLUTION CHANGE NEW LOT, BOT 850	5/25/20
2/02/20	19350	34.00 <sup>c</sup>	.09	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.10	LOWELL	WEEKLY SOLUTION	5/25/20
2/08/20	19350	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.10	34.00 <sup>c</sup>	.10	LOWELL	WEEKLY SOLUTION	5/25/20

SUPERVISOR TEST

ONEIDA CITY  
POLICE DEPT  
DATAMASTER dmt: 137806

Date: 02/08/2020

OPERATOR NAME:  
STEVEN J LOWELL  
Permit #: 40046  
Operator Agency/Dept: ONEIDA CITY PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:04
INTERNAL STANDARD	VERIFIED	20:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08

