

**CITY OF ONEIDA POLICE DEPARTMENT
CALIBRATION TEST RECORD**

SERIAL NUMBER 137806

NATIONAL PATENT ANALYTICAL SYSTEMS INC.
DATAMASTER DMT

DATE	LOT #	TEMP	TEST	TEMP	TEST	TEMP	TEST	TEMP	TEST	OPERATOR	COMMENTS	EXPIRES
12/16/19	19250	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	3/30/20
12/22/19	19250	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Mossack	WEEKLY SOLUTION	3/30/20
12/27/19	19250	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	Mossack	WEEKLY SOLUTION	3/30/20
12/31/19	19250	33.99c	.10	34.00c	.10	34.00c	.10	34.00c	.10	LOWELL	SOLUTION CHANGE BOT #923	3/30/20
01/03/20	19250	34.00c	.10	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	3/30/20
01/9/20	19250	34.00c	.09	34.00c	.10	34.00c	.10	34.00c	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/15/20	19250	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/21/20	19250	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/27/20	19250	34.00c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	3/30/20
1/31/20	19350	34.00c	.10	34.00c	.10	34.00c	.10	34.00c	.10	LOWELL	SOLUTION CHANGE NEW LOT, BOT 850	5/25/20
2/02/20	19350	34.00c	.09	34.00c	.10	34.00c	.10	34.00c	.10	LOWELL	WEEKLY SOLUTION	5/25/20
2/08/20	19350	34.00c	.10	34.00c	.10	34.00c	.10	34.00c	.10	LOWELL	WEEKLY SOLUTION	5/25/20
02/16/20	19350	33.99c	.09	34.00c	.09	34.00c	.09	34.00c	.09	LOWELL	WEEKLY SOLUTION	5/25/20

SUPERVISOR TEST

ONEIDA CITY
POLICE DEPT
DATAMASTER dmt: 137806

Date: 02/16/2020

OPERATOR NAME:
STEVEN J LOWELL
Permit #: 40046
Operator Agency/Dept: ONEIDA CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07

