

SUPERVISOR TEST

ONEIDA CITY
POLICE DEPT
DATAMASTER dmt: 137806

Date: 02/20/2020

OPERATOR NAME:
MATTHEW J MOSACK
Permit #: 45770
Operator Agency/Dept: ONEIDA CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:29
INTERNAL STANDARD	VERIFIED	23:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:30
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:31
BLANK TEST	0.00	23:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:32
BLANK TEST	0.00	23:33

